

CONSTRUCTION TO BE DONE BY FACILITY:

- 32" MINIMUM OPENING, 61" MINIMUM HALLWAY SWING CLEARANCE
- MOUNTING SUPPORT BEHIND THE 3' X 3' 60LB TREADMILL CONTROL PANEL
- ALL APPROVALS, PERMITS, AND AGENCY SUBMITTALS
- WET ENVIRONMENT WITH SPLASHING, DESIGN ROOM ACCORDINGLY
- AIR VENT 50CFM MINIMUM IN EACH ROOM WHERE EQUIPMENT IS LOCATED
- FLOOR BELOW TREADMILL FOOT PRINT TO BE LEVEL WITHIN .5" PER 7FT


- ELECTRICAL TO BE DONE BY FACILITY:**
- 230V | 1PH | 60HZ | 50AMP | 4 WIRE ELECTRICAL SUPPLY WITH A SIEMENS GFI 50-AMP SPA PANEL, **MODEL # W0408L1125SPA50 (REQUIRED)**
 - SPA PANEL MUST BE WITHIN 10FT OR LESS OF TREADMILL CONTROL PANEL
 - EQUIPMENT WILL RUN BELOW 230VAC, BUT SHORTENS LIFE EXPECTANCY AND VOIDS WARRANTY (AN AQUABOOST TRANSFORMER P/N0789490 RAISES 20VAC)
 - MOUNT CONTROL PANEL AND CONNECT INCOMING POWER PER NEC CODE
 - AFTER INSTALL BOND ALL EQUIPMENT PER NEC CODE

- PURCHASED ITEMS:**
- 0792001 AQUAFIT
 - 500 GALLON TANK w/RECIRC.
 - 0790573 500 GALLON TANK STAND
 - 0790892 EXTERIOR HAND RAILS AT ENTRY

- PLUMBING TO BE DONE BY FACILITY:**
- 3" FLOOR DRAIN NEAR TREADMILL ENTRY RECOMMENDED
 - 12" X 12" RECESSED FLOOR DRAIN NEAR EQUIPMENT REQUIRED
 - COLD WATER SUPPLY WITH HOSE BIB, FOR FILLING WATER STORAGE TANK
 - HOT WATER WITH HOSE BIB FOR MAKEUP, TREADMILL & CARTRIDGE CLEANING

- HUDSON SUPPLIED:**
- CUSTOM DESIGNED LAYOUT DRAWING WITH THE PURCHASE OF PRODUCT
 - INSTALLATION OF PURCHASED ITEMS AS PER FINAL SIGNED DRAWING
 - ON SITE IN-SERVICE TRAINING PROVIDED AFTER INSTALL
 - ELECTRICAL CONNECTIONS FROM PANEL TO HEATER, PUMPS, MOTOR, & HMI
 - "ASK SAM" WATER CARE SERVICE & AFTER HOURS SERVICE SUPPORT
 - PROPOSED PLUMBING PATH FROM TREADMILL TO EQUIPMENT
 - 3 FT DRIVE SHAFT CLEARANCE AREA FOR TREADMILL SERVICE

FACILITY LAYOUT:



Note: Splashing of water out of treadmill can occur, make sure this wall is water resistant.

--- EXAMPLE DRAWING SHOWN, CALL ENGINEERING FOR A CUSTOM LAYOUT ---

DRAWN BY: HJS	SALESMAN: XXXXX XXXXX	REV. DATE: 4-2-24	DRAWING #: XX-XXXX	FACILITY NAME: AQUAFIT PRO SAMPLE (SPACE SAVER STAND) LAYOUT DRAWING
------------------	--------------------------	----------------------	-----------------------	---

THIS DRAWING AND INFORMATION CONTAINED THEREON ARE CONFIDENTIAL AND THE PROPERTY OF HUDSON AQUATIC SYSTEMS, LLC. THEY ARE NOT TO BE REPRODUCED, DISTRIBUTED, DISCLOSED OR USED FOR ANY PURPOSE WITHOUT THE EXPRESS WRITTEN CONSENT OF HUDSON AQUATIC SYSTEMS.

--- FINAL DRAWING SIGN OFF ---
 PRODUCT TO BE BUILT AND INSTALLED AS SHOWN:
 APPROVED BY: _____
 DATE: _____

WEIGHTS:

TREADMILL EMPTY	1,250 LBS
TREADMILL w/48" WATER	5,400 LBS
500 GALLON TANK EMPTY	215 LBS
500 GALLON TANK w/WATER	4,500 LBS

FLOOR LOADS:

TREADMILL w/48" WATER	302 LBS PER SQ. FT.
500 GALLON TANK w/WATER	354 LBS PER SQ. FT.